




America

# CERTIFICATE

No. U8 095914 0002 Rev. 01

**Holder of Certificate:** **Amico Mobility Solutions**  
 122 East Beaver Creek Road, Building B  
 Richmond Hill ON L4B 4V1  
 CANADA

**Certification Mark:**



**Product:** **General Medical Devices**  
**Patient lift**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

**Test report no.:** 7169000014-200

**Date,** 2020-10-29

( Peter Keith )



America

# CERTIFICATE

No. U8 095914 0002 Rev. 01

**Model(s):** GoLift 400, GoLift 700  
**Brand Name:** Amico

**Tested according to:** CAN/CSA-C22.2 No. 60601-1:2014 excluding Biocompatibility (clause 11.7), Usability (clauses 12.2 and 15.1), PEMS (clause 14) and EMC (clause 17)  
 ANSI.AAMI ES60601-1:2005/A1:2012 excluding clauses 11.7 (Biocompatibility), usability (clauses 12.2 & 15.1) and 17 (EMC)

**Production Facility(ies):** 095914

## Parameters

Rated input voltage: 100-240Vac  
 Rated input frequency: 50-60Hz  
 Rated input current: 2.0-1.0A  
 Protection class: I  
 Degree of protection: Type B Applied Part