America

CERTIFICATE

No. U8 095914 0003 Rev. 00

Holder of Certificate: **Amico Mobility Solutions**
55 East Wilmot Street
Richmond Hill ON L4B 1A3
CANADA

Production Facility(ies): 095914

Certification Mark:

Product: **General Medical Devices**
Patient lift

Brand Name: **Amico**

Model(s): **GoLift Portable 450**

Parameters:

Device:
- Rated Input Voltage: 36 Vdc
- Rated Input Current: 1.11A
- Ingress Rating: IP23
- Degree of protection: Type BF applied parts

Charger:
- Rated Input Voltage: 100-240 V~
- Rated Frequency: 50-60 Hz
- Rated Input Current: 2.0 - 1.0 A
- Protection Class: II

Tested according to:

CAN/CSA-C22.2 No. 60601-1:2014-03 excl. Biocompatibility (Cl. 11.7), PEMS (Cl. 14), and EMC (Cl. 17).
ANSI/AAMI ES60601-1:2005/A1:2012-08 excl Biocompatibility (Cl. 11.7), PEMS (Cl. 14) and EMC (Cl. 17)
CAN/CSA-C22.2 No. 60601-1-11:2015
CAN/CSA-C22.2 No. 60601-1-6:2011/A1:2015-10

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 7169005130-000

Date, 2019-07-08

(Peter Keith)

TÜV SÜD America Inc. • 10 Centennial Drive • Peabody • MA 01960 • USA

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT
TÜV SÜD